

Interactive comment on “Investigation of water adsorption and hygroscopicity of atmospheric particles using a commercial vapor sorption analyzer” by Wenjun Gu et al.

Y.-H. Zhang

yhz@bit.edu.cn

Received and published: 30 July 2017

P15, Line 292-296
Figure 4 The authors mentioned that the measured mass hygroscopic growth factors agree well with the E-AIM model for $(\text{NH}_4)_2\text{SO}_4$ and NaCl at temperature of 25 and 5 degree centigrade. According to Figure 4, the data of NaCl measured agree well with model. But in Figure 4 (a) and (b), it seems that the $(\text{NH}_4)_2\text{SO}_4$ data from this work are obviously higher than that from the model, conflicting with the author’s description.

Interactive comment on Atmos. Meas. Tech. Discuss., doi:10.5194/amt-2017-56, 2017.

Printer-friendly version

Discussion paper

