

Atmos. Meas. Tech. Discuss., referee comment RC2  
<https://doi.org/10.5194/amt-2021-270-RC2>, 2021  
© Author(s) 2021. This work is distributed under  
the Creative Commons Attribution 4.0 License.

## **Comment on amt-2021-270**

Anonymous Referee #1

---

Referee comment on "The Aerosol Research Observation Station (AEROS)" by Karin Ardon-Dryer et al., Atmos. Meas. Tech. Discuss., <https://doi.org/10.5194/amt-2021-270-RC2>, 2021

---

Referee #1

Apologies, there was a typo in my review (RC1).

Page 5/Line 148: According to the manual of the 11-D monitor, the number concentration can reach up to 3 000 000 particles/L (= 3 000 particles/cm<sup>3</sup>, not 3 000 000 particles/cm<sup>3</sup> as stated in the manuscript) and mass concentration up to 100 mg/m<sup>3</sup> (not 100 000 µg cm<sup>-3</sup> as stated in the manuscript).