

Atmos. Meas. Tech. Discuss., author comment AC1
<https://doi.org/10.5194/amt-2021-425-AC1>, 2022
© Author(s) 2022. This work is distributed under
the Creative Commons Attribution 4.0 License.

Reply on RC1

Brayden Nilson et al.

Author comment on "Development and evaluation of correction models for a low-cost fine particulate matter monitor" by Brayden Nilson et al., Atmos. Meas. Tech. Discuss.,
<https://doi.org/10.5194/amt-2021-425-AC1>, 2022

Thank you for the time you put into reviewing our manuscript and the very useful and helpful feedback which has led to improvements in the paper. Please see our following responses and proposed alterations which we believe will resolve your individual comments.

Comment 1:

Were the authors trying to develop simple models to correct ambient PM2.5 concentrations reported by PurpleAir monitors across a range of locations and seasons in Canada and the United States? If so, how does this work build upon and differ from that of Barkjohn et al. [DOI: 10.5194/amt-14-4617-2021]?

Response 1:

See response #2

This work differs from and builds upon that of Barkjohn et al. since we include comparisons of the model from Barkjohn plus several other sources, our selected model form (Model 2) is different, we evaluate on an hourly time scale (not 24-hour averages like Barkjohn et al.), we include Canadian sites, and we use the AQHI+ system.

Comment 2:

It sounds like the authors might have been more interested in developing correction models that would help PurpleAir monitors predict "high" and "very high" AQHI+ levels correctly and expected wildfire smoke to be the most common cause of high or very high AQHI+ levels in Canada and the US. If so, the manuscript would benefit from discussions of (a) the existing literature describing the PurpleAir response to wildfire smoke [for example, Delp and Singer, DOI: 10.3390/s20133683 and Holder et al., DOI: 10.3390/s20174796] and (b) how the PurpleAir response to smoke is affected by the physics governing the PMS5003 sensor operation. Results from multiple field studies indicated that PMS5003 sensors overestimate smoke concentrations, and recent work by Ouimette et al. [DOI: 10.5194/amt-2021-170] indicated that this overestimation might be due to the small size of the particles produced during combustion. It is unlikely that the PMS5003 overestimates PM concentrations associated with wildfire smoke because of hygroscopic growth alone, so I'm a bit concerned by the authors' conclusion that Model 2,

which was designed to account primarily for RH effects, is best for correcting PurpleAir data reported when PM2.5 concentrations are "high" or "very high". I guess the k-value that was fit empirically helps account for some of the other factors.

Response 2:

We never stated that the "high" and "very high" AQHI+ levels should be focused on for correction performance. We state multiple times (see lines 17, 99, 154, 301, 317) that the "moderate" to "high" levels should be focused on, given that health messaging does not change until >3 AQHI+ and when AQHI+ is greater than 10 the messaging is the same (air quality is poor). We say that models 2 and 7 consistently perform better in high and very high on line 269, but they also perform well below that (just some models perform as well or slightly better). As well as on line 275 we say model 1 performs worse than model 7 in the very high range, but we don't discuss model 1 much otherwise. Nowhere else do we mention "very high" concentrations aside from when describing AQHI+ categories.

In our revised document we made edits to our introduction and added the following sentence to more clearly state our study objectives.

"It is our objective to create a correction model for general application across multiple sensors/locations, however, a more specialised correction is recommended where nearby colocation data are available."

Comment 3:

Regardless of the authors' objectives, it would be great to see more discussion of the physics that govern PMS5003 sensor operation and parameters that can affect the accuracy of PM2.5 concentrations reported by these sensors.

Response 3:

We added more information on how the nephelometers operate to our paragraph in the introduction already discussing the PurpleAir monitors and potential errors.

"[...]; concentrations are derived by correlating this scattering amplitude with a mass-based monitor (Hagan and Kroll, 2020)."

Comment 4:

Considering that the relationship between the ambient PM2.5 concentration and the light scattering signal received by the PMS5003 depends on particle size distribution, shape, refractive index, and density, and that these parameters can vary with location, season, and day-to-day weather, why do the authors conclude that a simple correction such as Model 2 or Model 7 is likely to produce acceptable results across a wide range of locations and seasons? Are there data on ambient aerosol properties in Canada and the US that the authors can cite to support this argument?

Response 4:

We acknowledge that simple corrections will never be perfect in all circumstances given the limitations imposed by the hardware and the wide range of particle properties that can exist and that a nephelometer cannot differentiate; we state on line 322 that where possible a site specific correction model should be developed, this is meant to be a general use model that performs better on average at most locations. On line 317 we state that in the future a seasonal/location specific model should be developed.

Comment 5:

I appreciate simple correction approaches because they contribute to the goal of making PM2.5 monitors and monitoring data accessible to a wide range of stakeholders, but I also think readers will have reasons to be skeptical of these correction models, and the authors should acknowledge and discuss those concerns.

Response 5:

We added the following sentence warning about the concerns behind using these models generally.

"Caution is advised when utilizing generalised models such as these, as they will not provide the same degree of improvement at all locations given differences in aerosol properties that a nephelometer cannot detect or differentiate."

Comment 6:

The authors argue that it's most important for PurpleAir monitors to predict "high" and "very high" AQHI+ values correctly, but I'm not sure I agree. When PM2.5 concentrations are high or very high, it's usually clear to the public that air quality poses a health hazard. I think it might be more important for the PurpleAir to predict AQHI+ values between 2 and 6 correctly, so that they can alert the public to pollution levels that are elevated above the baseline AQHI+ of 1 but that individuals might not readily perceive as hazardous using their senses. This comment is just for the authors' consideration.

Response 6:

See response #2

Comment 7:

- Lines 72-73: "PM2.5 concentration is reported by the sensors using two different proprietary correction factors (PM2.5 CF 1 and PM2.5 CF ATM) which convert the estimated particle count in size bins into the reported concentrations." This statement is speculative at best and most likely incorrect. I suggest the authors delete it. As the authors describe on lines 58-62, the PMS5003 functions as a nephelometer and not an optical particle counter. Ouimette et al. [DOI: 10.5194/amt-2021-170] have shown that the sensor output labeled "number of particles with diameter beyond 0.3 µm in 0.1 L of air" is a measure of the amount of light scattered to the detector by particles passing through the sensor. The relationship between the amount of light scattered to the detector and the PM2.5 concentration reported by the sensor is unknown, but multiple published datasets indicate that the particle count distribution output by the PMS5003 is not accurate and largely invariant [see He et al., DOI: 10.1080/02786826.2019.1696015; Tryner et al., DOI: 10.1016/j.jaerosci.2020.105654; and Ouimette et al., DOI: 10.5194/amt-2021-170]. Tryner et al. show that PM1, PM2.5, and PM10 mass fractions calculated from the particle count data reported by the PMS5003 are not consistent with the PM1, PM2.5, and PM10 mass concentrations reported directly by the sensor [DOI: 10.1016/j.jaerosci.2020.105654, Figure 6]. Wallace et al. [DOI: 10.1016/j.atmosenv.2021.118432] also note this discrepancy in their Appendix ("Apparently, the hidden algorithms of the Plantower approach assign values to measurements that in some way depart from using the mass calculated from the numbers of particles in the size categories.").

Response 7:

Thank you for this helpful response and clarifying our misunderstanding of the operation of the Plantower sensor. In the revised document we clarified that PA reports particle count in bins as well as mass concentrations, both estimated from some unknown calibration applied to the scattering amplitude.

Comment 8:

- Lines 74-75: "The CF ATM correction factor is derived from Beijing atmospheric conditions while CF 1 was derived from a lab study using symmetrical particles of a known size and is recommended for use in industrial settings (Zhou, 2016)." Where did the authors obtain this information? The PMS5003 manual that the authors cite does not contain any such statements

Response 8:

The original citation (Zhou, 2016) states "CF=1□standard particle" and "[CF=atm,] under atmospheric environment" in the table in Appendix I.

We added "; Yang, personal communication, 2016" to the citation (Yang is from PlanTower, the manufacturer of the PMS5003 sensor, who we contacted for additional information when we initially began working with the PurpleAir monitors in 2016).

Comment 9:

- Lines 76-77: "A recent study has developed a particle count correction factor using US-based sites which shows promise however that was not tested here (Wallace et al., 2021)." I'm not sure why the authors chose to give this correction approach special attention in this paragraph. The approach of Wallace et al. wasn't tested by the authors of this study and approaches to which the authors did compare their data aren't discussed until Lines 91-97. I'm not convinced that the approach of Wallace et al. is promising, either, since it seems to be based on flawed assumptions: that the PMS5003 measures particle counts and that the PM2.5 concentrations reported by the PMS5003 are calculated from the particle count data (see Comment 1).

Response 9:

We removed the discussion of the Wallace et al. paper on lines 76-77 (and lines 313-315).

Comment 10:

- Lines 104-107: What was the domain for this analysis? Canada and the United States?

Response 10:

We added text stating that the domain covers Canada & USA.

Comment 11:

- Lines 132-133: Why not download hourly PM2.5 concentration observations in the US from EPA AQS? There is a several-month lag between when the data are collected and when the data are available in AQS, but data in AQS have been QC'ed.

Response 11:

We did not use the AQS data as our data flows were setup for Airnow already, which provides both Canadian and US data, and the Canadian monitors would not be available

from the EPA AQS.

Comment 12:

- Lines 137-138: "We further removed several sites after viewing scatter plots of their valid PA and FEM PM2.5 observations and determining the performance to be unsatisfactory relative to the other sites." What criterion/criteria was/were used to determine that performance was unsatisfactory? Pearson correlation < 50%? The answer is not clear based on the current phrasing of this sentence and the one that follows.

Response 12:

We replaced "determining the performance to be unsatisfactory relative to the other sites." with "observing a non-linear relationship quite different from other sites."

Comment 13:

- Lines 141-143: "PA RH values were restricted to the range 30%–70% (any values above/below this were set to 30% or 70%, respectively) as these values are near the efflorescence and deliquescence points typical of fine particulate matter (Parsons et al. 2004, Davis et al. 2015). Corrections utilizing RH tended to overcorrect observations at these extreme RH values." (a) Please specify the fraction of 1-hour average data points that were affected by this restriction. (b) What was the justification for modifying RH values below 30%? I don't recall seeing evidence in the literature that low RH values are a problem. Did the authors do a sensitivity analysis to see whether and how their results were affected by leaving in RH values below 30%? (c) It's unclear whether the authors decided to replace RH values above 70% based on prior results reported in the literature or based on the results of their own modeling. If this decision was based on prior results from the literature, the authors should cite relevant studies. If this decision was based on the authors' own modeling, it would be nice to see these results presented in the form of a sensitivity analysis.

Response 13:

- a) We specified the fraction of hours where RH was missing or outside of 30% - 70% right before Table 1 in the revised paper.
- b) We stated in the revised paper that the Barkjohn et al. study uses 24-hour average data which would tend to keep the average RH within the 30-70% range.

At extreme rh values the correction models can have an exponentially increasing impact on the raw data. We plotted all of our concentration data in a scatter (PA-FEM error on the y, RH on the x) and at RH outside of 30-70% the corrected PA error increased noticeably.

- c) We added a figure of all of our concentration data in a scatter (PA-FEM bias on the y, RH on the x) to the SI and discussed this in the revised paper at the end of section 2.2.

Comment 14:

- Line 145: "Solar radiation impacts were too difficult to estimate given the variations in siting at each of the locations." I don't think the lack of correction for solar radiation is a big concern. The temperature and RH reported by the PurpleAir are biased high and low, respectively, even when the PurpleAir is installed indoors, due to heat generated by the electronics, so adjusting for solar radiation would still not eliminate the bias.

Response 14:

This is correct, however, the heat from the electronics would be more consistent between sites and therefore easier to model and correct for. The fact that there are large temperature variations as well based on differences in siting makes it nearly impossible to correct for generally.

Comment 15:

- Lines 190-192: How did the authors assign sites to the training and testing datasets? Was this assignment done randomly? Or did the authors try to make sure the full range of geographic areas, climates, and seasons were represented in each dataset?

Response 15:

We added “Training/testing sites were randomly selected then adjusted (again randomly) to ensure representativeness across geographic areas and concentration ranges.” to this section in the revised document.

Comment 16:

- Line 206: “increases” should be “increased”.

Response 16:

We replaced “increases” with “increased” in the revised document.

Comment 17:

- Line 208: “The mean testing site concentrations had similar ranges for matching monitor types between the two data sets...” I don’t understand what the authors mean by this. First the authors refer to “testing site concentrations”, but then they refer to “the two data sets”. Are the two data sets testing and training? FEM and PurpleAir? What were the “matching monitor types”? Does this phrase refer to the type of FEM monitor (e.g., beta attenuation) or to FEM vs. PurpleAir?

Response 17:

We replaced “The mean testing site concentrations had similar ranges for matching monitor types between the two data sets” with “Both the FEM and PA monitors had similar ranges of site median concentrations between the testing and training data sets”

Comment 18:

- Figure 2: “Distributions of the Federal Equivalent Method (FEM) and PurpleAir training/testing sites median PM_{2.5} concentrations ($\mu\text{g m}^{-3}$) at dry (0%–33%), moderate (34%–66%) and humid (67%–100%) relative humidity (RH) groupings.” (a) How was each site categorised as dry, moderate, or humid? Based on the mean or median hourly RH at the site over the full data collection period? (b) Have the authors considered making the groups 0-30%, 30-70%, and 70-100% since they chose to modify RH values in the 0-30% and 70-100% ranges?

Response 18:

- a) Each site was not categorized by their RH profile. Instead, the median values were calculated for every site for each of the “dry”, “moderate”, and “wet” hours. For each site, the median of the “dry” hours was calculated along with that for the “moderate” and “wet” hours.

We added the following summary of % of hours in each group to make this clearer.

"For the training data, 23% of the hourly observations were classified as "Dry" ($RH \leq 33\%$), 56% were classified as "Moderate" ($30\% < RH < 70\%$), and 21% were classified as "Humid" ($RH \geq 70\%$). The testing data were similar with 25%, 55%, and 19% of the hours classified as dry/moderate/humid, respectively.

- b) We believe this small change would not significantly impact the results. We chose equal 33% groupings instead of 30%,40%,30%.

Comment 19:

- Figure 3: Why were hours where the FEM AQHI+ was equal to 1 removed? For what fraction of hours was the FEM AQHI+ equal to 1? Most of them, right, considering the medians in Figure 2?

Response 19:

We clarified in the caption and the preceding text that these concentrations make up the bulk of the observations (skewing the results) and are unimportant from a health and management perspective. See also summary of the observation data presented in the response to Referee #2, comment 1.

Comment 20:

- Table 2: Model 1 is similar to the model fit by Barkjohn et al., but with slightly different coefficients. Did the authors fit Model 1 to PurpleAir data that had been adjusted to replace RH values $< 30\%$ with 30% and RH values $> 70\%$ with 70%? If so, did the authors also try fitting Model 1 to a dataset without adjusted RH values? It would be interesting to see how much the coefficients fit using the dataset from this study differ from the coefficients fit by Barkjohn et al.
- Table 2: Did the authors test Model 7, which was fit by Barkjohn et al., using the dataset in which RH values $< 30\%$ had been replaced with 30% and RH values $> 70\%$ had been replaced with 70%? If so, I'm not sure that's a fair test of this model because it wasn't fit using such adjusted RH values.

Response 20:

Yes, all of our models were fit using the truncated RH data. This improved the fit of all our models utilizing RH. We have clarified this in the text by adding "Truncating the RH data to 30 % - 70 % consistently improved the performance of RH-based models." to section 2.3

See comment #13. Barkjohn et al. used 24-hour averages to fit their data which effectively truncates the RH to ~30-70 in most locations. Therefore Model 1 would be quite comparable to that from Barkjohn et al.

Comment 21:

- Figure 4: The color scheme used here was confusing to me. My initial reaction was that overestimates should be red and underestimates should be blue. Did the authors choose red for underestimates because they view the PurpleAir incorrectly underestimating the AQHI+ value, and therefore failing to alert the public to the true extent of the health hazard posed by air pollution, to be the worse outcome?

Response 21:

The colour scale (purple to orange) is not meant to display preference to bias either way, as low and high bias are both bad. The grey/white colour is the desired value. Orange and purple are easier to differentiate for most types of colour blindness (see "Accessible palettes" at <https://davidmathlogic.com/colorblind/>).

Comment 22:

- Lines 256-264: Did the authors consider interpreting their results using the performance targets proposed in the US EPA Air Sensor Guidebook [document ID EPA/600/R-14/159]? In Section 5, the guidebook suggests precision and bias error < 50% for educational and informational purposes, < 30% for hotspot identification and characterisation or personal exposure monitoring, and < 20% for supplemental monitoring.

Response 22:

This guidebook is outdated (released June 2014) - a more recent set of protocols was defined in EPA Document ID 350785 (*Performance Testing Protocols, Metrics, and Target Values for Fine Particulate Matter Air Sensors: Use in Ambient, Outdoor, Fixed Site, Non-Regulatory Supplemental and Informational Monitoring Applications*) released February 2021. Table 3-1 defines metrics to consider, but thresholds are not provided.

We added the following discussion of this to section 2.4.

"Duvall et al. (2021) outline several key metrics to consider for small sensor performance: precision, bias and error, linearity, effects of RH and temperature, sensor drift, and accuracy at high concentrations. Evaluating precision is not viable in this study given that many sites only had a single PA installed. We will evaluate bias, error, and linearity through our analysis, as well as the effects of RH. We found temperature impacts to be minimal for our dataset, especially when the impacts of RH were already considered. Sensor drift is outside of the scope of our study, and accuracy at high concentrations is less of a concern given our use of the AQHI+ scale and focusing on the moderate to high concentrations."

Comment 23:

- Lines 299-300: "The Canadian AQHI+ system was useful as a framework for evaluating correction models across a range of concentrations infrequent high values or numerous low values can skew performance statistics when evaluating the full range at once." Is there a word missing here? Was this supposed to be two sentences?

Response 23:

We inserted ", as" between "concentrations" and "infrequent".

Comment 24:

- Lines 301-304: The names Model 2, Model 7, etc. are not very informative to a reader who is not looking at Table 2. It would be helpful to describe the key features of Models 2 and 7 here. Please also explain that models using RH as a predictor were fit and evaluated after replacing RH values below 30% with 30% and RH values above 70% with 70%. I don't think it's necessary to note Models 3 through 6 in this paragraph.

Response 24:

We added clarification to this paragraph briefly describing each model. Ie. "... Model 7, the humidity multiple regression from Barkjohn et al.". We also removed discussion on Models 3 through 6 here.

Comment 25:

- Lines 305-306: "...the average performance across the testing sites and over time was evaluated here; performance at colocation sites and across time was not the same." Why did the authors choose to focus on the average performance across testing sites and over time? Were there any sites or times where Models 2 and 7 performed notably better or worse? If so, what were the notable features of these sites and times (weather, unique PM sources) and what do those features say about the advantages and limitations of Models 2 and 7?

Response 25:

It was our objective to "provide an optimised correction model for North American PA sensors" - without further data/effort a regional/seasonal specific model cannot be developed so a more general model was required, especially for utilizing the PA sensors to support AQ management.

We added a figure to the SI breaking down the mean AQHI+ bias figure into individual testing sites and referenced it at the start of section 3.2 as well as in section 4. Site 37 (in Oregon) had the most notable improvement after correcting, sites 43-45 (in Nebraska, Alaska, and California, respectively) had the most notable reduction in performance after correcting. All of these sites are in considerably different geographic locations - outlining the difficulty in developing a regional based correction model without additional colocation sites.

Comment 26:

- Lines 313-315: "...the improved particle bin correction factor proposed by Wallace et al. (2021) should be implemented for these sites..." See Comment 3. I don't understand why the authors repeatedly refer the correction approach proposed by Wallace et al. I don't think it's a good approach or particularly relevant to the work presented here.

Response 26:

See response # 9. Removed the Wallace et al. discussion.