

Atmos. Meas. Tech. Discuss., editor comment EC1  
<https://doi.org/10.5194/amt-2021-72-EC1>, 2021  
© Author(s) 2021. This work is distributed under  
the Creative Commons Attribution 4.0 License.

## Editor's comment

Mingjin Tang (Editor)

---

Editor comment on "A method for liquid spectrophotometric measurement of total and water-soluble iron and copper in ambient aerosols" by Yuhan Yang et al., Atmos. Meas. Tech. Discuss., <https://doi.org/10.5194/amt-2021-72-EC1>, 2021

---

In the conclusion, detection limits of Fe and Cu in the air are provided. I think the detection limits in air depends on the volume of air sampled. Should the authors also provide detection limits in the solutions?