

Magn. Reson. Discuss., referee comment RC1  
<https://doi.org/10.5194/mr-2021-11-RC1>, 2021  
© Author(s) 2021. This work is distributed under  
the Creative Commons Attribution 4.0 License.

## **Comment on mr-2021-11**

Anonymous Referee #1

---

Referee comment on "Overhauser dynamic nuclear polarization (ODNP)-enhanced two-dimensional proton NMR spectroscopy at low magnetic fields" by Timothy J. Keller and Thorsten Maly, Magn. Reson. Discuss., <https://doi.org/10.5194/mr-2021-11-RC1>, 2021

---

The authors present an important topic for Magnetic Resonance, the development of a new method in the area of 2D NMR-DNP to study liquid samples. They demonstrate clearly the newly developed approach for compact low field spectrometers. They solve a systematic problem, significant field drift, common with permanent magnets used in NMR spectrometers. Also, it is appreciated and important for the NMR community that their code is easily accessible. Finally, I recommend that the submitted paper be published with minor corrections. See below.

Possible corrections:

line 152: do you mean Tecmag, not Techmag, correct?

In Figure 4a, shouldn't the x-axis be labeled "J-coupling (Hz)"?

Suggested comments to address:

It seems from the technical details that 2D acquisitions presented took about 30 mins. What are the limitations of the proposed method? Could it be applied to non-monotonic drift? Could the method be applied for much longer acquisition times particularly when the samples are not neat compounds?

What applications would benefit most from such instrument/method? Compared to more traditional compact NMR spectrometers?

