

Magn. Reson. Discuss., author comment AC1
<https://doi.org/10.5194/mr-2021-56-AC1>, 2021
© Author(s) 2021. This work is distributed under
the Creative Commons Attribution 4.0 License.

Comment on mr-2021-56

Neil MacKinnon et al.

Author comment on "Selective excitation enables encoding and measurement of multiple diffusion parameters in a single experiment" by Neil MacKinnon et al., Magn. Reson. Discuss., <https://doi.org/10.5194/mr-2021-56-AC1>, 2021

We sincerely thank the community and reviewers for their time and critical feedback to our manuscript. We believe we have addressed all of the comments, and the manuscript has gained significantly in clarity. Below are our point-by-point responses, together with the adjustment that was made in the manuscript:

Reviewer 1 (Davy Sinnaeve):

- **The authors mention that wider segments of the spectrum can also be selected, encompassing multiple signals. But what if the signals within one segment are coupled to one other? Would that not generate artefacts from non-refocused J-coupling evolution?**

If several spins are selected by the selective pulses that are coupled to each other, coupling evolution indeed takes place during each selective encoding/decoding element. As a result, some in-phase signal will be transferred to antiphase, which will then be removed by the crusher gradient at the end of each selective element. As such, only a reduction in in-phase signal intensities will be experienced in the weak coupling limit. The signal intensities are reduced according to $\cos(\pi J t_{\text{eff}})$, where t_{eff} is given by the time of selected spins spent in the transverse plane, which depends on the gradient lengths, delays, and the specific selective pulses used in the sequence. For wider selected areas corresponding selective pulses are typically short, on the order of 1 or few milliseconds. As a consequence t_{eff} is also in the range of few ms and with typical small 1H-1H-coupling constants on the order of -15 to +10 Hz, effective signal reductions are on the order of 10% (assuming a coupling constant of 15 Hz and $t_{\text{eff}}=8\text{ms}$, the cos-dependence causes a 7% reduction in intensity). This can usually be tolerated.

If coupled spins experience severe second order artefacts, signal distortions might occur, but these distortions will be the same for each gradient strength, so that diffusion measurements should still be possible.

Manuscript adjustment: a statement noting the reduction in signal in case of J-coupling has been added to the discussion section.

- **The SNR for the Ala methyl signal increased significantly for the selective experiment compared to the non-selective experiment (Table 1). This is unexpected, given the expected increased relaxation losses during the selective pulses. Is there an explanation for this? Is the same observed for the Methionine methyl signal?**

After a closer look at the SNR script, we identified an error in the noise calculation. This has been corrected, results and discussion updated.

Manuscript adjustment: Recalculated SNR values Ala, Met, Water = -2, +4, -33% difference between selective and non-selective experiments.

- **The SNR and diffusion coefficient measured on the Methionine methyl group are not shown in Table 1, not discussed in the text, and the decay curves are not shown in Figure 3. Yet this was the third signal selected. It would be good to include this, to observe the consistency with the Ala signal.**

Manuscript adjustment: Figure 3 was updated.

- **Some remarks about the representation of the pulse sequence of Figure 1:**

Manuscript adjustment: the pulse sequence was updated taking into account all suggestions. The new version is a better representation of what was actually used in this work, and the figure caption notes where the flexibility in the sequence lies.

- **The Stejskal-Tanner equation (1) is valid for rectangular gradient shapes. The gradient shape is not mentioned in the experimental part. It would be good to confirm the gradient shape in the experimental part, since for Bruker these are standardly SINE or SMSQ shapes, requiring a slightly modified ST equation.**

Manuscript adjustment: We used SMSQ gradients, and is noted in the materials and methods section. A clarification statement is added declaring that no correction for slightly non-ideal rectangular pulses was applied in this work.

- **Also, it should maybe be explicitly noted that the value of τ is actually quite significant here to correctly take into account, since it also encompasses the long selective 180° pulse (14.7 ms). If different selective pulse durations would be used for each spectral segment (as is possible in the pulse sequence code provided on zenodo), the ST-equation will be significantly different for each segment.**

Manuscript adjustment: this clarification was added to the materials and methods section.

- **Figure 3a shows the signal intensity as a function of k, while Figure 3b shows it as a function of gradient strength percentage. This inconsistency seems unnecessary to me.**

Manuscript adjustment: Figure was updated.

- **On page 3, line 66, ledbpge2s1d should be ledbpgp2s1d.**

Manuscript adjustment: correction made.

Reviewer 2 (anonymous):

- **The exact benefit of using different gradient ramps for the different resonances should be clarified. Is it expected to improve the precision of the measured diffusion coefficient? The trueness? Both? This is hard to understand in the present manuscript. Maybe the authors should report the uncertainty of the fit for the selective and non-selective experiments? Also, in Table 1, the value of the diffusion coefficient for water measured with the selective experiment is further way from the calibration value than measured with the non-selective experiment. Please explain if that it an improvement, and why?**

We observed small deviations in the extracted diffusion coefficients between the selective and non-selective experiment variants. While a reduction in the error was observed for the extracted water diffusion coefficient in the optimized selective experiment, the benefit was small (data added to the manuscript). We aim to emphasize that the benefit of the selective experiment is measurement parallelization. As an example, the technique is expected to bring considerable advantages in speedup for DTI (Diffusion Tensor Imaging) applications, since the 3D diffusion tensor (with up to 6 components) needs to be determined on a voxel-by-voxel basis. In this case, the ability to encode up to 6 scalar components in a single k-space acquisition, i.e. in one shot and subsequent FID, would be a tremendous advantage. Another example is a restricted diffusion situation, where one can simultaneously measure with optimal diffusion parameters a fast species and a slow species experiencing restriction. A plot of simulated data is presented to highlight this example.

Manuscript adjustment: a new paragraph was added to the discussion. The anticipated gain in DTI and multi-component diffusion is discussed. A new figure was added: 3 plots of simulated diffusion curves demonstrating that in a single experiment, one can identify slow, multi-component diffusion at the same time as a fast, single component diffusing species.

- **Why does SNR increase in the selective experiment for the MAS signal? Is that a positive observation? Or is that a reflection of the fact that coherence transfer pathway selection is not as effective?**

After a closer look at the SNR script, we identified an error in the noise calculation. This has been corrected, results and discussion updated.

Manuscript adjustment: Recalculated SNR values Ala, Met, Water = -2, +4, -33% difference between selective and non-selective experiments.

- **Please provide the values of the encoding gradient in G/cm or T/m, or the maximum gradient that can be delivered by the probe. These probes typically deliver up to 200 or 300 G/cm. Together with the diffusion delay of 250 ms and a ramp of up to 95%, this would results in an unusually strong attenuation. Please clarify.**

The value of the gradient was determined to be 260 G / cm. The diffusion experiment

parameters were adjusted to observe the decay of MAS. With the same parameters, the water signal decayed with 40% of maximum gradient strength (Figure 3).

Manuscript adjustment: value of the gradient strength inserted into methods section.

- **The authors state that the experiment is accelerated by a factor of n . In the example that they report, $n = 3$ but two of the resonances belong to the same molecule, and are encoded with the same parameters. So, isn't the acceleration by a factor of 2 only?**

This is a valid point, here we demonstrate $n = 2$ since the parameter we adjusted was the gradient amplitude, which was the same for the two MAS signals. The general expected acceleration is still true, so long as the n encodings are unique.

Manuscript adjustment: clarification was added to the conclusion.

- **More generally, stating that the experiment is accelerate will only be valid once the authors convincingly explain why there is an advantage of using different ramps.**

Addressed in the related comments.

- **The author state that the approach is applicable to any diffusion experiment. This is not obvious, in particular for spin echo based experiment. Once one of the resonances is in the transverse plane, it experiences any subsequent gradient pulse. How would it work to encode independently the different resonances ?**

This is indeed true, we had incorrectly over-generalized the applicability of this concept. The method requires magnetization storage along the z-axis, and therefore is more correctly applicable to stimulated echo methods. This statement has been clarified in the manuscript.

Manuscript adjustment: clarification that the experiment can be generalized to stimulated echo experiments was implemented in the discussion and conclusion.

- **It would help to discuss examples where one would want to use this pulse sequence. If the gradient ramp are tailored for the selected resonances, it means that the corresponding diffusion coefficients are know at least approximately. So what would the accelerated experiment be used for? The introduction covers a broad range of applications of diffusion NMR, but no connection is made, in the discussion, between these area of application and the proposed method. For example, the abstract states the possibility to encode multiple delay or multiple directions; does that assume that several resonances are available for the same molecule?**

The discussion has been updated to highlight the applicability (also in response to an earlier comment), in particular in connection with the concepts from the introduction.

Manuscript adjustment: new paragraph was added to the discussion.

- **It would be helpful to also show the pulse sequence without the loop structure, that is, the complete pulse sequence with a fixed value of n. Please also indicate, in this figure, the exact definition of Delta. This would help the reader to see that the Delta delay is the same for all of the selected resonances.**

The pulse sequence diagram has been updated accounting for all reviewer suggestions.

Manuscript adjustment: Figure 1 updated.

- **Why use two different steps to encode the two MAS resonances, rather than a single one using a dual-band selective pulse?**

We agree that using multi-band pulses can be used in this experiment, which was originally argued as a means to effectively access shorter diffusion times since less time is spent in selective encoding (last sentence of the discussion section). In this work, it was our object to demonstrate the 3 signals could be selectively treated, even if 2 of the signals originated from the same molecule.

Manuscript adjustment: the last sentence of the discussion was updated, exchanging "multi-selectivity" with "multi-band selectivity" to clarify this point.

Community comment (R. Soong)

- **Have the author try to use this measurement in liquid crystal environment to evaluate anisotropic diffusion tensors?** We do have preliminary data of a selective diffusion experiment in a PBLG sample. In the experiment, three signals originating from a single molecule free to diffuse within the aligned sample were exposed to X, Y, and Z diffusion encoding / decoding gradients selectively, in a single experiment (i.e. measurement of diffusion in 3 directions in one experiment). Our preliminary data suggest there was an observable difference in the diffusion behavior between the Z and X,Y diffusion directions. We require significantly more time and experiments to confirm this result, and hope to report in the near future.
- **How realistic to use this sequence to extract diffusion tensor and map out the orientation of its environment.**

We believe this could be one of the key experiments standing to gain in terms of measurement time from the selective diffusion experiment. In principle, 6 separate measurements are required to define the diffusion tensor, and in practice many more directions are often measured. By increasing the number of directions measured in a single experiment by taking advantage of the selective encoding, the measurement time required to define the tensor should significantly improve.